

Laboratory Accreditation Programmes

Schedule to
CERTIFICATE OF ACCREDITATION

| | | | | | | | |
|---|---|------------------|------------|----------------------|------------|--------------------|------------|
| Hill Laboratories Wellington | Client No. 9558 | | | | | | |
| Level 3, 204 Thorndon Quay, Pipitea, Wellington, 6011 | | | | | | | |
| Telephone 04 472-1030 | www.hill-laboratories.com | | | | | | |
| Authorised Representative Mrs Debra McLachlan Wellington Quality Assurance Coordinator PO Box 16607 Hornby Christchurch 8441 New Zealand | | | | | | | |
| Programme Chemical Testing Laboratory | | | | | | | |
| Accreditation Number 1301 | Initial Accreditation Date 18 May 2018 | | | | | | |
| Conformance Standard NZS ISO/IEC 17025:2005 General requirements for the competence of testing and calibration laboratories | | | | | | | |
| Testing Services Summary 2.64 Asbestos | | | | | | | |
| Key Technical Personnel <table> <tr> <td>Mr Henry Collins</td> <td>2.64(a)(b)</td> </tr> <tr> <td>Mr Dexter Paguirigan</td> <td>2.64(a)(b)</td> </tr> <tr> <td>Mr Vikram Pathania</td> <td>2.64(a)(b)</td> </tr> </table> | | Mr Henry Collins | 2.64(a)(b) | Mr Dexter Paguirigan | 2.64(a)(b) | Mr Vikram Pathania | 2.64(a)(b) |
| Mr Henry Collins | 2.64(a)(b) | | | | | | |
| Mr Dexter Paguirigan | 2.64(a)(b) | | | | | | |
| Mr Vikram Pathania | 2.64(a)(b) | | | | | | |

| | | |
|--|-----------------------------|-------------|
| Authorised: General Manager <i>P. Bam</i> | Issue 3 Date: 07/02/19 | Page 1 of 2 |
|--|-----------------------------|-------------|

Schedule to
CERTIFICATE OF ACCREDITATIONHill Laboratories
Chemical Testing Laboratory
SCOPE OF ACCREDITATION

Accreditation Number 1301


2.64 Asbestos**(a) Identification**

Bulk identification of asbestos using polarised light microscopy (PLM)

AS 4964:2004 Method for the Qualitative Identification of Asbestos in Bulk Samples

(b) Fibre Counts**Analysis**

Determination of air-borne asbestos fibre counts on membrane filters

National Occupational Health & Safety Commission: 3003(2005) Guidance Note on the Membrane Filter Method for Estimating Airborne Asbestos Fibres (2nd Ed)Authorised:
General Manager

Issue 3

Date: 07/02/19

Page 2 of 2