


Laboratory Accreditation Programmes

Schedule to	
CERTIFICATE OF ACCREDITATION	
Hill Laboratories Microbiology - Hamilton	Client No. 4064
Private Bag 3205, Waikato Mail Centre, Hamilton, 3240 1 Clow Place, Melville, Hamilton, 3206	
Telephone 0508 445-5522 Fax	www.hill-laboratories.com
Authorised Representative Mrs Shobhna Ram Microbiology Team Leader	
Programme Drinking Water Testing Laboratory	
Accreditation Number 845	Initial Accreditation Date 30 January 2003
Conformance Standard NZS ISO/IEC 17025:2005 General requirements for the competence of testing and calibration laboratories Drinking-Water Standards for New Zealand 2005 (Revised 2008)	
Testing Services Summary	
1.12	Waters (NZ Drinking Water Standards - Microbiology)
Key Technical Personnel	
Mrs Varsha Boby	1.12
Mrs Shobhna Ram	1.12 (selected)
Mr Kevin Wang	1.12

Authorised: General Manager		Issue 38	Date: 10/02/17	Page 1 of 2
--------------------------------	---	----------	----------------	-------------

Schedule to
CERTIFICATE OF ACCREDITATION

Hill Laboratories
 Drinking Water Testing Laboratory
SCOPE OF ACCREDITATION

Accreditation Number 845

The following scope of accreditation provides for the testing of potable waters for the purpose of assessing compliance with the Drinking-Water Standards for New Zealand 2005 (Revised 2008)


1.12 Waters (NZ Drinking Water Standards - Microbiology)

(a) Potable waters

In accordance with APHA “Standard Methods for the Examination of Water and Wastewater” (22nd Edition) except where otherwise indicated

Determinand	Method Reference	WINZ Method Code
Total coliforms - MPN	9221 B	MTF
Total coliforms - Colilert P/A	9223 B	COLILE
Total coliforms - Colilert Quantitray	9223 B	ColiQT
Faecal coliforms - MPN	9221 E	MTF
Faecal coliforms - MF	9222 D	MF
<i>E coli</i> - MPN	9221 F	MTF
<i>E coli</i> - Colilert P/A	9223 B	COLILE
<i>E coli</i> - Colilert Quantitray	9223 B	ColiQT

Authorised:
 General Manager



Issue 38

Date: 10/02/17

Page 2 of 2